

# ST-120 Sound Level Calibrator



IEC 60942:2003 Class1/ANSI S1.40-1984.

- Output sound level : at 94dB and 114dB.
- Frequency : 1KHZ  $\pm 1\%$ , Class1, A
- Accuracy :  $94 \pm 0.3\text{dB}$   $114 \pm 0.5\text{dB}$

Specifications

- Sound pressure levels : at 94dB and 114dB.
- For 1/2" condenser microphone calibration.
- Low battery indication.
- Harmonic waves distortion  $\pm 3\%$ .
- Output feedback detection.
- Power: 9V battery 006P or IEC6F22 or NEDA 1604
- Operation environments :  $23.0\text{oC} \pm 3\text{oC}$   
 $< 85\%RH$  ,850mbar~1050mbar,  $\leq 2000M$